

Batch Certificate

Certificate Number : 2507/139C/8878
 Product Code : FlawGlo SWP 139
 Batch Number : 8878
 Date of Manufacture : June 2025
 Date of Expiry : July 2027

Parameters	Required		Remarks
Corrosiveness	Non-corrosive	-	Acceptable
Removability	Satisfactory with solvent	-	Acceptable
Viscosity	5.63cst	-	Acceptable
Appearance	Dark red oily liquid	-	Acceptable
Drying time	Non-drying	-	Acceptable
Wettability	Satisfactory	-	Acceptable
Contaminants	% by weight of material	% by weight of residue of material	Remark
Sulfur	< 0.1 %	< 0.01 %	Acceptable
Halogen (Cl + F)	< 0.1 %	< 0.01 %	Acceptable

Performance test: Detect 50 micron & 30-micron cracks on Ni-Cr Panel & on Aluminium quench crack block when used with FlawGlo Developer & Cleaner.

We hereby certify that the above batch Meets the requirements of -

- A) ISO 3452-2 Standard
- B) ASTM E165 -12
- C) ASME BPVC Section V

We further certify that this material does not contain mercury as a basic element and no mercury bearing equipment was used in its manufacture.

For Quality Control



Product Executive - Chemicals